



January 22, 2016

Mrs. Alicia Czerniawski  
ANRIC ENTERPRISES INC.  
701 Evans Ave, Suite 202  
Toronto, ON, M9C 1A3

RECEIVED

FEB 03 2016

SC

**Type of device:** CORIOLIS FLOW METERS CAT-'F'

**Manufacturer:** AW-LAKE COMPANY  
8809 Industrial Dr.  
Franksville, WI  
U.S.A. 53126

Dear Madam,

According to our law respecting pressure vessels, your design has been reviewed and registered as follows:

**CRN N°: 0F18057.56**

Our reference No: 956809

Main Design No. : 150327-TCM-7900\_CRN\_WB REV.B

Note: If the CRN is a five digit number following the letter and start with a 0, the 0 is not considered to be an integral part of the CRN.

Yours truly,

LK

For:

Moussa Mekhnach

Inspecteur autorisé (e)  
Direction de l'inspection de la fabrication  
d'appareils sous pression

## Statutory Declaration Registration of Fittings

**(a) Design Qualification**

I<sup>1</sup> Mark Iverson  
(Name of applicant)  
General Manager  
(Position eg, president, plant manager, chief eng.)  
of AW-Lake Company  
(name of company)  
Located at 8809 Industrial Dr. Franksville, WI 53126  
(plant address)

do solemnly declare that the fittings listed hereunder, which are subject to the Boilers & Pressure Vessels Act:

- comply with all the requirements of the ANSI/ASME codes as to their dimensions, material, identification & service for which are required:  
Or  
 are not covered by the provisions of the ANSI/ASME codes, and are therefore constructed to comply with ASME B31.3, 2012 Edition code and standard, and are designed to the best current engineering practice, as shown by the supporting test data.

**(b) Quality control of Manufacture**

I further declare the manufacture of these fittings is controlled by a quality control program which complies with the requirements of ISO 9001:2008, and has been verified by the following authority or authorized agency Intertek

The fittings<sup>2</sup> covered by this declaration, for which I seek registration, are Coriolis Flow Meters CAT-'F'

In support of the application, the following information, calculations and/or test data are attached:  
Scope of Registration - (150914-TCM-CRN-SOR), Product Catalogues & Design Drawings

Declared before me at 3421 16<sup>th</sup> Street

In the of County of Kenosha, State of Wisconsin

The 22 day of September AD 2015

[Signature]  
A (commissioner for oaths)

[Signature]  
Signature of Declarer

**QUEBEC**

For Official Use Only	
The application is accepted for registration in Category <u>F</u> in accordance with the Boilers and Pressure Vessels Act and CSA Standard B51.	<b>ENREGISTRÉ</b> EN VERTU DE LA LOI A.S.P. CHAP. A 20.01 <span style="font-size: 2em; font-weight: bold;">OF 18057.56</span>
This registration must be revalidated after ten (10) years from the date of acceptance.	
Registered Number CRN _____	For the Chief Inspector _____ Date _____ SIGN: <u>[Signature]</u>

DATE: 21 JAN. 2016

<sup>1</sup> Three completed copies of Statutory Declaration form together with three copies of Catalogs, drawings of Bulletins illustrating above fittings shall be submitted.  
<sup>2</sup> All fittings are required to be registered in the name of the Manufacturer.  
<sup>3</sup> This form shall be completed and signed by the president of highest official in the manufacturing plant where the fitting is produced.

21 JAN. 2016




## Scope of Registration

Model & Size	Design Pressure	Design Temperature	Materials of Construction
TCM 65k - 2"	1450 psi	302F	Tubes: A 269 Type 316L Splitter: Cast A 351 CF3M
TCM 28k - 1"	1450 psi	302F	Tubes: A 269 Type 316L & Hastelloy B622- N06022 Splitter: Cast A 351 CF3M & Hastelloy A 494, CX-2MW
TCM 5500 - 1/2"	5000 psi	302F	Tubes: A 269 Type 316L Splitter: Cast A 351 CF3M
TCM 7900	1450 psi	302F	Tubes: A 269 Type 316L Splitter: Cast A 351 CF3M
TCM 325	2900 psi	302F	Tubes: A 269 Type 316L Splitter: Machined 316L/1.4404
TCM 650	2900 psi	302F	Tubes: A 269 Type 316L Splitter: Machined 316L/1.4404
TCM 3100	2900 psi	302F	Tubes: A 269 Type 316L Splitter: Machined 316L/1.4404
TCM 1550	2900 psi	302F	Tubes: A 269 Type 316L Splitter: Machined 316L/1.4404

Name & Position: Mark Iverson General Manager

Signature & Date:

 9/14/10

Flow Measurement Technology from TASI Flow Global Brands

