



the pressure equipment safety authority

9410 - 20 Ave N.W.

Edmonton, Alberta, Canada T6N 0A4

Tel: (780) 437-9100 / Fax: (780) 437-7787

January 05, 2016

**Attention:** Alicia Czerniawski  
ANRIC ENTERPRISES INC  
701 EVANS AVENUE STE 202  
TORONTO, ON M9C 1A3

RECEIVED  
30 JAN 14 2016

The design submission, tracking number 2015-08768, originally received on December 18, 2015 was surveyed and accepted for registration as follows:

**CRN :** 0F18057.52

**Accepted on:** January 05, 2016

**Reg Type:** New Design

**Expiry Date:** November 12, 2025

**Drawing No. :** FLOW METERS TCM-CR-SOR

**Fitting type:**

Design registered in the name of : AW LAKE COMPANY

**The registration is conditional on your compliance with the following notes:**

*This registration is valid only for fittings fabricated at the location(s) covered by the QC certificate attached to the accepted AB-41 Statutory Declaration form. This registration is valid only until the indicated expiry date only if the Manufacturer maintains a valid quality management system approved by an acceptable third-party agency until that date. Should the approval of the quality management system lapse before the expiry date indicated above, this registration shall become void.*

An invoice covering survey and registration fees will be forwarded from our Revenue Accounts.

Enclosed are stamped prints for your reference.

Sincerely,


GRYNCHUK, MILLA



the pressure equipment safety authority

STATUTORY DECLARATION  
Registration of Fittings

In this space, show facsimile of manufacturer's logo or trademark as it will appear on the fitting.



I, Mark Iverson,

General Manager

(company title, e.g. vice president, plant manager, chief engineer) (must be in a position of authority)

of AW-Lake Company  
(name of manufacturer)

located at 8809 Industrial Dr. Franksville, WI 53126  
(plant address)

do solemnly declare that the fittings listed hereunder, which are subject to the Safety Codes Act (check one)

comply with the requirements of \_\_\_\_\_ which specifies the dimensions,  
(title of recognized North American Standard)

are not covered by the provisions of a recognized North American standard and are therefore manufactured to comply with ASME B31.3, 2012 Edition as supported by the attached data which identifies the dimensions, materials of construction, pressure/temperature ratings and the basis for such ratings, and the marking of the fittings for identification.

I further declare that the manufacture of these fittings is controlled by a quality control program which has been verified by the following authority, Intertek as being suitable for the manufacture of these fittings to the stated standard. The fittings covered by this declaration, for which I seek registration, are Coriolis Flow Meters CAT-'F'

In support of this application, the following information, calculations and/or test data are attached:

Scope of Registration - (150914-TCM-CRN-SOR), Product Catalogues & Design Drawings

DECLARED before me at 3421 16<sup>th</sup> in the City of Kenosha State of Wisconsin  
this 17 day of September, 2015  
(Month) (Year)

(print) Atty. H. R. Iverson

(sign) [Signature]  
(A Commissioner for Oaths)

[Signature]  
(Signature of Applicant)

For Office Use Only

To the best of my knowledge and belief, the application meets the requirements of the Safety Codes Act and CSA Standard B51, Clause 4.2, and is accepted for registration in Category F

Registration Number: OF 18057.52

Date Registered: JAN 05 2016

[Signature]  
(For the Administrator/Chief Inspector of Alberta)

Expiry Date: 2025-11-12

**ADKA**  
SAFETY CODES ACT - PROVINCE OF ALBERTA

**REGISTRATION OF FITTINGS**

REGISTRATION NO. **OF 18057.52**

**AW-LAKE**  
COMPANY

DWG. NO. or CAT. NO. Scope **Scope of Registration**

TYPE OF FITTINGS Flow meters  
**Model & Size**  
**Design Pressure**  
Date **JAN 05 2016**  
INITIALS  
MILLA GRYNCHUK, P.Eng.  
DESIGN SURVEY ENGINEER

Model & Size	Design Pressure	Design Temperature	Materials of Construction
TCM 65k - 2"	1450 psi	302F	Tubes: A 269 Type 316L Splitter: Cast A 351 CF3M
TCM 28k - 1"	1450 psi	302F	Tubes: A 269 Type 316L & Hastelloy B622- N06022 Splitter: Cast A 351 CF3M & Hastelloy A 494, CX-2MW
TCM 5500 - 1/2"	5000 psi	302F	Tubes: A 269 Type 316L Splitter: Cast A 351 CF3M
TCM 7900	1450 psi	302F	Tubes: A 269 Type 316L Splitter: Cast A 351 CF3M
TCM 325	2900 psi	302F	Tubes: A 269 Type 316L Splitter: Machined 316L/1.4404
TCM 650	2900 psi	302F	Tubes: A 269 Type 316L Splitter: Machined 316L/1.4404
TCM 3100	2900 psi	302F	Tubes: A 269 Type 316L Splitter: Machined 316L/1.4404
TCM 1550	2900 psi	302F	Tubes: A 269 Type 316L Splitter: Machined 316L/1.4404

**Name & Position:** Mark Iverson General Manager

**Signature & Date:** *Mark Iverson* 9/14/10

Flow Measurement Technology from TASI Flow Global Brands

